

PLEASE READ THE MANUAL BEFORE USING!



THE MANUAL IS AVAILABLE FROM PRIOR SCIENTIFIC AND SHOULD BE READ AND UNDERSTOOD BEFORE INSTALLATION AND USE. THE MANUAL IS AVAILABLE FROM WWW.PRIOR.COM

- · Adhere to all warnings on the product and in the user manual.
- This product MUST NOT be exposed to open flames extremely hot or cold temperature, water or moisture. DO NOT allow objects to fall, or liquids to spill, on the product.
- The ProScan III is a Class 1 system and must be grounded.



- Connect the AC power cord only to designated power sources as marked on the product. Make sure the electrical cord is located so that it will not be subject to damage.
- Ensure the connection to the power source and the front panel power switch are easily accessible.
- Do not connect or disconnect anything to or from the ProScan III unless the controller is turned off first.
- DANGER Never alter the AC cord or plug. The power cord set must be an appropriately rated and approved cord set in accordance in the regulations of the country it is used in. If the supplied plug adapter is not the correct fitting for your geographic area or if you are unsure about the relevant regulations, please contact your supplier for advice.
- Do not attempt to disassemble, tamper or repair this product; it contains NO user serviceable parts and doing so may void the warranty. Service should be performed ONLY by authorized service centers.
- Clean only the exterior of this product using a damp, lint free cloth.
- Installing older revision firmware is never recommended and may render unit inoperable.
- Please contact Prior for information on updating controller firmware if deemed necessary. New units ship with the latest firmware and should not need updating.
- Please contact Prior Scientific at 781-878-8442 or an approved representative if you have any questions or comments.

CONNECTIONS

ALWAYS MAKE SURE THE POWER TO THE PROSCAN III IS OFF BEFORE CONNECTING OR DISCONNECTING ANY COMPONENTS



Scientifi

PROSCAN[®]III QUICK SET UP


ALWAYS MAKE SURE THE POWER TO THE PROSCAN III IS OFF BEFORE CONNECTING OR DISCONNECTING ANY COMPONENTS Ensure the correct encoder cable goes into the correct port. Read manual before use (www.prior.com).

٢


0

٢

Control



Connection between ProScan III and Prior device




Stage

6

ill, Hy

0

Illumination



____ Encoded connection between ProScan III and encoded Prior device